



Product Service

CERTIFICATE

No. Z2 095104 0025 Rev. 00

Holder of Certificate: Ocean Insight Inc.

3500 Quadrangle Blvd
Orlando FL 32817
USA

Certification Mark:



Product:

Electrical equ. for measurement, control and laboratory use
Near Infrared Spectrometer

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition, the certification holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the testing and certification regulations of TÜV SÜD Group have to be complied. For details see: www.tuvsud.com/ps-cert

Test report no.: 72186512-000

Valid until: 2028-08-16

Date, 2023-08-29

(David Dorfner)



CERTIFICATE

No. Z2 095104 0025 Rev. 00

Model(s): NR

Brand Name: Ocean Insight or OceanOptics

Parameters:

Rated Input Voltage:	5 VDC
Rated Input Current:	3.5 A
Protection Class:	Not directly connected to mains
Ambient Temperature:	40°C max

Tested according to: EN 61010-1:2010/A1:2019
BS EN 61010-1:2010/A1:2019